

Information and Technical Data

Fully automated measurement

with preconfigured parameters tailored to your glove type and detection requirements

Instant ready for measurement

just unpack, insert and start the automated measurement

Automatic glove filling

no additional air supply needed with automatic glove filling pump

Superior hole detection

of holes smaller than 100 µm while maintaining short measurement times

Fast glove integrity verification

typically commencing from 8 to 15 minutes total test time, depending on glove material

Reliable in-situ testing

without jeopardizing glove integrity by glove removal or manipulation

100% mobility

lightweight and completely wireless glove testing system

Customizable to your requirements

in size and shape, even in form factor and material

Cleanroom suitable

IP64 certification and easily replaceable filter for air intake

Easy cleaning and sterilization

glove tester build with minimal grooves and edges

Status-related information

pressure gauge, battery status, measurement information and results shown on digital display

Simplicity of operation

intuitive two button interface

cGMP compliant testing

Long operating time

for repetitive measurements to validate your entire isolator/RABS without troubles

Electronically controlled clamping and sealing of glove port

with pneumatic inflated rubber seal

Slim and ergonomic design

thin casing of only 5 cm incl. automatic glove filling pump

Services

Custom glove integrity test parameters to your requirements

Qualification Services (IQ, OQ and more)

Regular service and maintenance of glove testers and equipment



Legal Foundations

DIN EN ISO 14644-7

VDI 2083 Part 16.1

ISO 1779:1999

GAMP 5

21 CFR Part 11

EU Annex 11

Adherence to standards and instructions in respect to security issues

Glove Testing Devices are developed according to DIN / ISO / EN standards

Accident prevention regulations (UVV regulations)

2006/42/EG

2004/108/EC

2006/95/EC

EN 55011:2011-0

EN 61000-4-2:2009-12

EN 61000-4-3:2013-04

EN 60335-2-29:2010-11

IP 64 for glove testing devices

